



Nova Scotia Health Authority

2019 Radiation Safety - Cyclotron Public Disclosure Program

Summary Report

The NSHA Radiation Safety Program provides information to the public about radiation safety in the cyclotron department, located at the VG site of the QEII Health Sciences Centre, as a requirement of the organization's licence.

Section 1: General

- There were no reportable radiation safety incidents in 2019.

Section 2: Personnel Dosimetry

- No staff member reached a radiation dose exposure limit.

Section 3: Cyclotron Facility

- The facility is operating within the licenced parameters.
- The annual radiation safety audit found no major issues of compliance.
- Annual radiation monitoring surveys were completed with no issues identified.
- The annual compliance report was submitted and accepted by the regulators.

To complete our public engagement survey or for more information, please visit the following link or contact the Radiation Safety Office.



<https://www.engage4health.ca/cyclotron-radiation-safety-public-information-disclosure-program>

Radiation Safety Office
QEII Bethune Building, Room 244
1276 South Park Street, Halifax, Nova Scotia
B3H 2Y9
902-473-2767